Examination in ultraviolet light

When examination in ultraviolet light is required, an ultraviolet lamp having a maximum output of approximately 254 nm or 365 nm is to be used.